



Govt. of Assam
OFFICE OF THE PRINCIPAL
JORHAT INSTITUTE OF SCIENCE & TECHNOLOGY
JORHAT – 785010, Assam
E-mail: principaljistnew@gmail.com

No.JIST/Tender/2021-22/31

Date-07/01/2022

Quotation Invitation

Selected quotations affixing court stamp Rs. 8.25 only in plain papers are hereby invited by the undersigned for supply of **Analytical Balance** for Chemistry Department.

The filled quotations along with all the required document must reach in the office of the Principal, Jorhat Institute of Science and Technology, Jorhat on or before 21st January 2022 at 12 noon and will be opened on the same date at 3p.m. in the office chamber of the undersigned.

The firm who quoted the tender will be eligible for supply of items in future for the financial year 2021-22 and 2022-23 as per lowest rate approved by this office.

The envelope containing the quotation would please be sealed and super-scribed as under.

“Quotation for supply of Analytical Balance for Chemistry Department”

Terms and conditions

- 1) The quotation received after this deadline & unsealed shall not be entertained under any circumstances. In case of postal delay this Institute will not be responsible. The offer submitted Fax/Email shall not be considered, and no correspondence will be entertained in this matter.
- 2) Quotation must on the letter head of the firm duly signed by the proprietor/Partner/Director of their representative, in case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation.
- 3) Rates must be quoted in Indian rupees and as per the format specified taxes extra if any must be written separately.
- 4) No overwriting or cutting is permitted in the rate. If found the quotation shall be summarily rejected.
- 5) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
 - a) Firm shall be registered with the Government of Assam/Central Government.
 - b) The firm shall have valid GST/Other taxes and IT PAN.
 - c) The firm should not be black listed by any Govt. Agency/Dept.
- 6) Payment Terms: Payment will be only after satisfactorily delivery/ commissioning of material.
- 7) Bidder must quote the product as per specification provided in annexure-I.

Principal (i/c)


Jorhat Institute of Science and Technology

Jorhat-10
Principal

Jorhat Institute of Science & Technology
Jorhat-10, Assam

Technical Specification of analytical balance

Capacity	Up to 300g
Readability	0.0001g (0.1mg)
Repeatability (\pm)	0.0001g
Linearity (\pm)	$\pm 0.0003g$
Min. (Capacity)	0.01g
Display	LCD
Pan Size	90 mm \emptyset
Response Time	4-6 Sec
Calibration	Internal Calibration
Tare Range	Full
Operating Temperature	+15°C To +30°C
Humidity	50-70% RH
Warm up time	30 min.
Unit	g, ct, ozt, oz, GN, dwt, lb
Key Sound	Yes
Weighing Chamber Height	210mm
Data Output Mode	Command and continuous output
Power Supply	AC 220V/50-60 Hz



PRINCIPAL
JORHAT INSTITUTE OF SCIENCE & TECHNOLOGY
JORHAT-10